

Product datasheet

AAV6 ELISA Control

Short overview

Cat. No. PRAAV6-C

Quantity 1 vial (1.0E+09-1.6E+09 capsids)

Concentration 2.0E+09-3.3E+09 capsids/ml after reconstitution in 500 µl ASSB 1x (please find the lot-specific

concentration on the CoA and on the vial)

Product description

Formulation Lyophilized, empty AAV6 capsids. Reconstitute in 500 µl ASSB 1x (provided with PROGEN's

AAV6 Titration ELISA), incubate 5 min at RT and mix by rolling 5 min. Avoid vortexing! The final

solution contains stabilizing protein, phenol red and ASSB 1x buffer.

Stability 4 weeks at 2-8°C after reconstitution in ASSB 1x

Storage 2-8°C

Intended use Research use only

Application ELISA

Applications

ELISA As a positive control in ELISA, a 1:4 dilution in ASSB 1x (provided

with PROGEN's AAV6 Titration ELISA) and analysis at least in

duplicates is recommended.

Background

The AAV6 ELISA Control consists of fully assembled, empty AAV6 capsids. The concentration is lot specific.

The AAV6 ELISA Control can be used as a positive control with PROGEN's AAV6 Titration ELISA (PRAAV6). If you require different lots of the AAV6 ELISA Control and the Kit Control included in your PROGEN AAV6 Titration ELISA, please enquire to check availability.

Reading of the AAV6 ELISA Control from the standard curve (i.e. Kit Control, included in PROGEN's AAV6 Titration ELISA) is influenced by interand intra-assay, but also by inter-lab variances. Therefore, it is recommended that each laboratory determines its own acceptable range of recovery.

The AAV6 ELISA Control has been calibrated on an internally established reference standard. The internal reference standard is a preparation of full AAV6 capsids that has been characterized by qPCR and ddPCR (DNA quantification) and electron microscopy (ratio of full to empty capsids). For further information, please see as an example our poster The New AAV3 Titration ELISA.

Product images



AAV6 ELISA Control